

CERTIFICATE OF ACCREDITATION

In terms of section 22(2) (b) of the Accreditation for Conformity Assessment, Calibration and Good Laboratory Practice Act, 2006 (Act 19 of 2006), read with sections 23(1), (2) and (3) of the said Act, I hereby certify that:-

MONTECH CALIBRATION SERVICES (PTY) LTD
Co. Reg. No.: 2014/182132/07
PRESSURE CALIBRATION LABORATORY

Accreditation Number: **CAL 071-02-00**

is a South African National Accreditation System accredited Calibration laboratory provided that all SANAS conditions and requirements are complied with

This certificate is valid as per the scope as stated in the accompanying scope of accreditation Annexure "A", bearing the above accreditation number for

PRESSURE METROLOGY

The facility is accredited in accordance with the recognised International Standard

ISO/IEC 17025:2017

The accreditation demonstrates technical competency for a defined scope and the operation of a laboratory quality management system

While this certificate remains valid, the Accredited Facility named above is authorised to use the relevant SANAS accreditation symbol to issue facility reports and/or certificates

Mr F Osman
Acting Chief Executive Officer

Effective Date: 10 March 2026
Certificate Expires: 09 March 2031

ANNEXURE A

SCOPE OF ACCREDITATION

PRESSURE METROLOGY

Accreditation Number: CAL 071-02-00

| <p>Permanent Address of Laboratory: Montech Calibration Services (Pty) Ltd Pressure Calibration Laboratory 77 Fabriek Street Strydompark Randburg 2169</p> <p>Postal Address: Postnet Suite 266 Privte Bag X21 Bryanston 2021</p> <p>Tel: (011) 464-5071 Cell: 072 779-4076 E-mail: info@moncal.co.za seola@moncal.co.za</p> | | <p>Technical Signatory: Mrs S Targett</p> <p>Nominated Representative: Mrs S Targett</p> <p>Issue No.: 05 Date of Issue: 10 March 2026 Expiry Date: 09 March 2031</p> | | |
|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------|-----------------------------------------------------------------------|
| ITEM | MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT | NOMINAL RANGE OF MEASURED QUANTITY | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | METHOD / PROCEDURE |
| 3 | PRESSURE | | | |
| 3.2 | Gauge Pressure | | | |
| 3.2.1 | Gas Medium <ul style="list-style-type: none"> • Pressure Gauge • Digital Manometer • Pressure Transmitters | - 80 kPa to 0 kPa 0 kPa to 2 000 kPa | 0,07 % + 10 Pa | Calibration against a reference pressure transducer and/or indicator. |
| 3.2.2 | Liquid Medium <ul style="list-style-type: none"> • Pressure Gauge • Digital Manometer • Pressure Transmitters | 0 Mpa to 70 Mpa 70 Mpa to 413 Mpa | 0,07 % per reading 0,8 % reading | Calibration against a reference pressure transducer and/or indicator. |
| 4 | On site calibration for item 3.2 above | | | |

Original Date of Accreditation: 20 January 2022

Page 1 of 1

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%

ISSUED BY THE SOUTH AFRICAN NATIONAL ACCREDITATION SYSTEM

Accreditation Manager